



August 7, 2009

## ELECTRONIC REPORT

Brooke M. Westlake  
Sales Manager & Sales Support  
Computerized Screening, Inc.  
9550 Gateway Drive  
Reno, Nevada 89521 USA

Dear Ms. Westlake,

Please find attached herewith a report entitled, "CSI HEALTH STATION MODEL 6K-BLOOD PRESSURE KIOSK VALIDATION STUDY: Comparison to Auscultatory Method (AAMI/ANSI SP-10)".

This validation study was initially conducted with the instrument described as the CSI Health Station 6K. I have examined the current product you have developed as the CSI Managed Health System (5K) and the advanced model, CSI Managed Health System (9K). The 5K and the 9K units employ electronics for the determination of blood pressure that are identical to the 6K instrument and thus, I conclude that the products known as the CSI Managed Health System (5K), and the CSI Managed Health System (9K) are substantially equivalent to the CSI Health Station 6K, and meet or exceed the AAMI/ANSI SP-10 standard (4.4.5.1.2.B).

We are pleased to be able to provide this report. Please do not hesitate to contact us with any questions about our report.

Regards,

A handwritten signature in black ink, appearing to read "Iain L. O. Buxton".

Iain L. O. Buxton  
Professor, University of Nevada School of Medicine  
Department of Pharmacology  
Principal, Aptus Consulting